



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:*

***East Tennessee Scale Works, Inc.  
10000 Martel Road  
Lenoir City, TN 37772***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025:2005***

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):*

***On-Site Calibration of Weighing Devices  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

*Initial Accreditation Date:*  
April 23, 2010

*Accreditation No.:*  
67070

*Issue Date:*  
April 23, 2010

*Certificate No.:*  
L10-56

*Expiration Date:*  
April 22, 2012

*Page No.:*  
Page 1 of 2



# Certificate of Accreditation: Supplement

**East Tennessee Scale Works, Inc.**

10000 Martel Road  
Lenoir City, TN 37772

*Accreditation is granted to this facility to perform the following calibrations:*

## Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	REMARKS
Scale and Balances	1 mg to 5 kg	$(1.20 \times 10^{-3} + 1.15 \times 10^{-4}Wt)$ g	Class F Weights NIST Handbook 44
	0.001 lb to 1 000 lb	$(1.16 \times 10^{-2} + 1.04 \times 10^{-4}Wt)$ lb	
	1 000 lb to 10 000 lb	$(1.16 \times 10^1 + 1.05 \times 10^{-4}Wt)$ lb	
	10 000 lb to 200 000 lb	$(23.0931 + 4.48 \times 10^{-5}Wt)$ lb	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.